Configura post-delivery checklist

Use this form to ensure the optimum set up of the Configura for your client after delivery. Please note the references to the relevant page in the Configura User Manual.

1) Is the user sitting at the back of the chair?	YES	NO
Possible adjustments: - Improve transfer method - Increase/decrease seat depth using Depth Adjustment kit (page 6) - Adjust filling in Multi-adjustable Pillow Backrest (page 7)	☐	☐
Comments		

2) Is the seat depth correct? (thigh supported in upright sitting)	YES	NO
Possible adjustments: - Increase/decrease seat depth using Depth Adjustment kit (page 6) - Adjust filling in Multi-adjustable Pillow Backrest (page 7)	☐	☐
Comments		

3) Is the seat height correct? (calf sufficiently supported when in recline)	YES	NO
Possible adjustments: - Increase/decrease seat height using Height Adjustment kit (page 5)	☐	☐
Comments		

4) Does the user have a good midline sitting position?	YES	NO
Possible adjustments: - Adjust filling in Multi-adjustable Pillow Backrest (page 7) - Try Lateral Support Backrest option (page 7)	☐	☐
Comments		

5) Does the user feel comfortable in the chair?	YES	NO
Possible adjustments: - Adjust filling in Multi-adjustable Pillow Backrest (page 7) - Try alternative seat base cushions (page 6) - Try alternative backrest (page 7) - Check transfer and improve if necessary	☐	☐
Comments		

6) Is the hand control within easy reach?	YES	NO
Possible adjustments: - Pull handset further through the handset loop - Change the side of the handset (page 9)	☐	☐
Comments		

7) Can the user operate the chair to rise and recline?	YES	NO
Possible adjustments: - Pull handset further through the handset loop - Change the side of the hand control (page 9)	☐	☐
Comments		

8) Can the user transfer in an acceptable fashion?	YES	NO
Possible adjustments: - Adjust rise setting (Tilt, Rise & Tilt, Vertical Rise) (page 10) - Increase/decrease seat height (page 5)	☐	☐
Comments		

9) Is there enough space for the chair to recline?	YES	NO
Possible adjustments: - Move chair forward from the wall	☐	☐
Comments		

10) Is the cushion fitted correctly in the seat base? (especially in the case of the Cushionair)	YES	NO
Possible adjustments: - Check Cushionair cushion is correctly fitted (cushion packer, blue cells, overlay and connect to pump) (page 6) - Consider whether the overlay should fit over the cushion or underneath the cushion (over the cushion is preferable to keep the cushion components together) (page 6)	☐	☐
Comments		